



FIGURE 1. Diagram showing the different pressure regimes in the modified Q-TOF 2 mass spectrometer (Micromass UK) (Sobott et al. 2002). Reducing the efficiency of pumping in the source and intermediate pressure regions, as well as admitting a collision gas to the collision cell, improves the transmission of high-mass species through the instrument.

Protein-Protein Interactions: A Molecular Cloning Manual, 2nd Ed., © 2005 by Cold Spring Harbor Laboratory Press, Chapter 17, Figure 1.